

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100684

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/27/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.17: 0.075
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 350 psi

BLANK TEST	0.000	09:24
INTERNAL STANDARD	VERIFIED	09:25
EXTERNAL STANDARD	0.070	09:25
BLANK TEST	0.000	09:26
EXTERNAL STANDARD	0.070	09:26
BLANK TEST	0.000	09:27
EXTERNAL STANDARD	0.071	09:27
BLANK TEST	0.000	09:28
EXTERNAL STANDARD	0.073	09:28
BLANK TEST	0.000	09:29
EXTERNAL STANDARD	0.070	09:29
BLANK TEST	0.000	09:30

Average = 0.0708
Std Dev = 0.0013

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C	PASSED
Breath Tube = 47.8°C	PASSED

PUMP INFO

Flow Rate = 3.861 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only: Instrument being removed from
service. BMB 3/8/18