

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100414

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/16/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.71: 0.079
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 1071 psi

BLANK TEST	0.000	12:23
INTERNAL STANDARD	VERIFIED	12:23
EXTERNAL STANDARD	0.079	12:23
BLANK TEST	0.000	12:24
EXTERNAL STANDARD	0.079	12:25
BLANK TEST	0.000	12:25
EXTERNAL STANDARD	0.078	12:26
BLANK TEST	0.000	12:27
EXTERNAL STANDARD	0.079	12:27
BLANK TEST	0.000	12:28
EXTERNAL STANDARD	0.079	12:28

User Abort

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 48.1°C	PASSED

PUMP INFO

Flow Rate = 4.971 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only DM 10/11/24