



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 1233
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 02511080A2
Expiration: 2/1/2013

0.080 BAC (For use with breath alcohol testing instruments)

Contents: 105 Liters @ 1025 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	208.4 ppm	+/- 0.002 or 2%	NDIR
Nitrogen	Balance	BAC whichever is greater	

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

MARCH 01, 2011

OPERATOR'S NAME:
ILN
OPERATOR'S NUMBER: 7732
ALCO TARGET VALUE: 0.080
ALCO S/N: 8092

— SUPERVISOR NAME —

.080 ADJUSTED FOR 30.06 in		
ALCO TARGET	.080	14:16
BLANK TEST	.000	14:18
INTERNAL STANDARD	VERIFIED	14:18
ALCO TV 30.06 in	.062	14:18
BLANK TEST	.000	14:19
ALCO TV 30.06 in	.061	14:19
BLANK TEST	.000	14:20
ALCO TV 30.06 in	.061	14:20
BLANK TEST	.000	14:21
ALCO TV 30.06 in	.062	14:21
BLANK TEST	.000	14:22
ALCO TV 30.06 in	.062	14:22
BLANK TEST	.000	14:23

N = 5
SIM. = .0003
AVG. = .0616
S.D. = .0005
S.E. = 1.5147


Specialty Gas Lab Manager

1-28-2011
Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

