

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100422

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 02/12/2020

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C  
Breath Tube = 47.5°C

PASSED  
PASSED

Detector Overflow