

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100667

Received From: APD

Date: 11/7/23

Analyst: Carlie Blaister

Reason for instrument return:

Instrument not recognizing when a breath sample is being provided.

Change instrument and ~~case~~ location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax *no case CWB 11/7/23*

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)