

# Alaska Scientific Crime Detection Laboratory

## DataMaster Calibration Form

Issued: 4/12/2018  
Effective: 4/13/2018

Version: DMCAF 2018 R0  
Status: Active

Instrument # 100691

Analyst Brandi Barnett

Date: 10/12/21

### Reason for Calibration

Recalibrate due to filter wheel maintenance.

### 1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Lifeloc
- c. Lot 21DP 41158
- d. Expiration Date 4/29/23

### 2. Wet Bath Simulator

- a. Serial Number DR5298
- b. Date Calibrated 6/14/21

### 3. Attach Calibration Factors

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100691  
-----

Date: 10/12/2021  
Time: 14:17:18

OPERATOR NAME:  
BRANDI BARNETT  
CERTIFICATE #: 7038

Ca = 0.1000  
CAL = 0.977643    0.800 <= CAL < 1.200  
b1 = 0.0000    0.0000 <= b1 < 0.0040  
b2 = 0.0057    0.0010 <= b2 < 0.0100  
b3 = 0.0000    0.0000 <= b3 < 0.0040  
Xq = 0.1030    0.0500 <= Xq < 0.2000  
a21 = 1.197996    1.050 <= a21 < 1.300  
a31 = 0.440779    0.300 <= a31 < 0.800