

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 2/9/2023

Instrument #: 100668

Instrument Location: Anchorage Jail – South processing room

Supervisor Name and Agency: Jason Cusack, Anchorage Police

Describe the Status Change or Issue with the Instrument:

I took this instrument out of service at the request of the crime lab.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 2/9/2023 0343

For Use by SCDL

Additional Notes

Instrument pulled for Filter Wheel errors. TLM 02/09/2023

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor