

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 7/17/15 Instrument Serial Number: 100695

Instrument Location: KAKE

Supervisor Name and Agency: TRP. UMBS / AST JUNEAU

Describe the Status Change or Issue with the Instrument:

ISSUES WITH EXTERNAL STANDARD BEING ERRATIC DURING VDC'S AND ALSO DURING SUBJECT TESTING. LAB REQUESTED DATAMASTER BE EXCHANGED.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 7/17/15 - 1005

For Use by SCDL

Additional Notes

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740