

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form – DMCAF2011R1		
	Issuing Authority – Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100415

Date: 6/24/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Repro
- c. Lot 10003
- d. Expiration Date 9/8/12

2. Wet Bath Simulator

- a. Serial Number DR4161
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100415

Date: ~~06/02/2011~~ 6/24/11
Time: 07:25:35 ~~06/24/11~~

SUPERVISOR NAME:
COLLEEN S OBRYANT
CERTIFICATE #: 6237

LOT #: REPCO100
BOTTLE #: 10003
EXPIRATION: 09/08/2012

Ca = 0.1000
CAL = 1.001237 0.800 <= CAL < 1.200
b1 = 0.0006 0.0000 <= b1 < 0.0040
b2 = 0.0044 0.0010 <= b2 < 0.0100
b3 = 0.0000 0.0000 <= b3 < 0.0040
Xq = 0.0970 0.0500 <= Xq < 0.2000
a21 = 1.190898 1.050 <= a21 < 1.300
a31 = 0.428205 0.300 <= a31 < 0.800

COB 2/2

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100415\B-100415 ECF-20140

2/18/2014