


| | | | |
|---|---|--------------------------|----------------------|
|  | State of Alaska Scientific Crime Detection Laboratory | | |
| | DataMaster Calibration Form - DMCAF2011R1 | | |
| | Issuing Authority - Nita Bolz, Scientific Director | | |
| | Approval date: 03/10/11 | Effective date: 03/10/11 | Archive date: Active |

DataMaster DMT Calibration Form

Instrument # ~~100~~ ¹⁰⁰ 100398Date: 5/18/11

1. Wet Bath Simulator Solution

- Level 0.100
- Manufacturer Guth
- Lot 10260
- Expiration Date 10/25/12
10/25/10

2. Wet Bath Simulator

- Serial Number DR-1161
- Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst

C. Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100398

Date: 05/18/2011
Time: 15:01:06

SUPERVISOR NAME:
COLLEEN S OBRYANT
CERTIFICATE #: 6237

LOT #: GUTEL00
BOTTLE #: 10260
EXPIRATION: 10/25/2018
exp 2
cod 5/19/11

| | | | | |
|-----|---|----------|--------|----------------|
| Ca | = | 0.1000 | | |
| CAL | = | 0.984202 | 0.800 | <= CAL < 1.200 |
| b1 | = | 0.0000 | 0.0000 | <= b1 < 0.0040 |
| b2 | = | 0.0077 | 0.0010 | <= b2 < 0.0100 |
| b3 | = | 0.0007 | 0.0000 | <= b3 < 0.0040 |
| Xq | = | 0.0909 | 0.0500 | <= Xq < 0.2000 |
| a21 | = | 1.228106 | 1.050 | <= a21 < 1.300 |
| a31 | = | 0.486055 | 0.300 | <= a31 < 0.800 |

CS 2/2

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100398\B-100398_ECF-20140

2/19/2014