

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 5-11-21 Instrument Serial Number: 100381

Instrument Location: Soldotna AST

Supervisor Name and Agency: Ronny Simmons / AST

Describe the Status Change or Issue with the Instrument:

New datamaster sent to Post from lab. A vac was done prior to taking the datamaster out of service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 5-11-21 @ 0958 hrs

For Use by SCDL

Additional Notes

Taken out of service due to high prevalence of EPI status messages. See in field instrument review 2021.02.22 - 2021.04.11
all 6/11/2021
JM 06/11/21

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740