

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100689

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/08/2020

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.56: 0.079
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 377 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 04:55 |
| INTERNAL STANDARD | VERIFIED | 04:55 |
| EXTERNAL STANDARD | 0.076 | 04:55 |
| BLANK TEST | 0.000 | 04:56 |
| EXTERNAL STANDARD | 0.077 | 04:56 |

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

| | |
|-------------------------|--------|
| Sample Chamber = 48.9°C | PASSED |
| Breath Tube = 40.5°C | PASSED |

PUMP INFO

| | |
|-----------------------|--------|
| Flow Rate = 4.190 L/M | PASSED |
|-----------------------|--------|

DETECTOR INFO

| | |
|----------|--------|
| PUMP ON | PASSED |
| PUMP OFF | PASSED |

FILTER INFO

| | |
|----------|--------|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

| | |
|-------------------|--------|
| INTERNAL STANDARD | PASSED |
|-------------------|--------|