

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 14 Dec 15 Instrument Serial Number: 100383

Instrument Location: 354SFS

Supervisor Name and Agency: SSgt Jacobs / 354SFS

Describe the Status Change or Issue with the Instrument:

Communication Error SC Temp: 0.0°C

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 14 Dec 15 / 1625

For Use by SCDL

Additional Notes

On 12/10/15, Joseph Jacobs called regarding Communication Error message displayed where Ready Push Run usually is. Also, sample chamber temperature and tank pressure were reading as zero. While still on the phone, I asked him to open the DMT and ensure that all lines were fully connected. This did not resolve the issue.
- CRF 12/15/15

Email completed form to dps.cdli@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100383

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/08/2015

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.01

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 0.0°C

FAILED

Breath Tube = 0.0°C

FAILED

Suck Back Error

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.

(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director

State Breath Alcohol Program

Subscribed and sworn before me this ____ day of _____, 20 ____