



State of Alaska
 Department of Public Safety
Scientific Crime Detection Laboratory

4805 Dr. Martin Luther King Jr. Ave.
 Anchorage, Alaska 99507
 David Kanaris, Chief, Forensic Laboratories

Breath Alcohol Calibration Report
Calibration Item: DataMaster DMT
Serial Number: 100673

Report Issue Date: 11/07/2023
Date Received: 2023-06-06
Condition of Item: Good

Method Used: DataMaster DMT Calibration / Current Breath Alcohol Procedure Manual
Calibration Range: 0.020 g/210L - 0.300 g/210L

Start Date 06/06/2023 - As Found - **End Date 06/07/2023** As Found - Evaluation Analyst: Tatiana Malnoske
 As Found Conditions - Barometric Pressure (inHg): 29.73 Barometer Used: Fisher S/N 221250180 Temperature (C): 22.5
 Relative Humidity (%): 32%

As-Found Data									
Nominal Values (g/210L)	Target Values (g/210L)	Lot Number	Results (g/210L)					Mean (g/210L)	
0.020	0.020	AG217902	0.020	0.020	0.020	0.020	0.020	0.020	0.020
0.080	0.080	09523080A1	0.080	0.080	0.080	0.081	0.080	0.080	0.080
0.300	0.298	AG217903	0.302	0.302	0.302	0.302	0.302	0.302	0.302

The internal standard value on the first ran VOC was 2.84% which is high, so I ran a second VOC and the value from that was 3.39% which is over the acceptable value.

I ran another VOC the next day, 06/07/2023, so ensure that the values from 06/06/2023 wasn't a fluke and the value from that VOC was 2.73% which is still high.

Start Date 10/31/2023 - As Left - **End Date 11/02/2023**

As Left Conditions - Barometric Pressure (inHg): 29.83 Barometer Used: Fisher S/N 221250180 Temperature (C): 21.4 Relative Humidity (%): 24

As-Left Data										
Nominal Values (g/210L)	Target Values (g/210L)	Lot Number	Results (g/210L)					Mean (g/210L)	Uncertainty (g/210L)	
0.020	0.020	AG217902	0.020	0.020	0.020	0.020	0.020	0.020	+/- 0.002	
0.040	0.040	AG217904	0.040	0.040	0.040	0.040	0.039	0.040	+/- 0.003	
0.080	0.080	AG310901	0.079	0.080	0.079	0.080	0.079	0.079	+/- 0.004	
0.100	0.100	AG211003	0.100	0.100	0.100	0.100	0.099	0.100	+/- 0.004	
0.150	0.150	AG134205	0.149	0.149	0.150	0.150	0.150	0.150	+/- 0.005	
0.300	0.299	AG217903	0.300	0.300	0.301	0.301	0.301	0.301	+/- 0.011	

Calibration is performed using dry gas standards traceable to NIST. Standards are sourced from an external provider accredited to ISO/IEC 17025 and ISO 17034 standards. Internal instrument barometer readings are traceable to NIST. External barometers are calibrated by an external calibration provider accredited to the ISO/IEC 17025 standard.

The As-Left results listed above include the estimated measurement uncertainty with a coverage factor k=2 and an expanded coverage probability of 95%. The estimated uncertainty applies to the calibration process and not to subject breath samples.

This report applies only to the item being calibrated and shall not be reproduced, except in full, without written approval from the laboratory.



State of Alaska
Department of Public Safety
Scientific Crime Detection Laboratory

4805 Dr. Martin Luther King Jr. Ave.
Anchorage, Alaska 99507
David Kanaris, Chief, Forensic Laboratories

Serial Number: 100673

Results authorized by Carlie Glaister on 11/02/2023

carlie.glaister@alaska.gov

Technically and administratively reviewed by Colleen O'Bryant on 11/07/2023

Alaska Scientific Crime Detection Laboratory

DataMaster Calibration Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100673 Analyst: Carlie Glaister Start Date: 10/31/23

- Diagnostic Test (attached)
- Adjust voltages, if needed
 - Lamp (2.60 V Maximum)
 - Bias
 - Cooler (2.00 V Maximum)
- Internal Standard value \leq 3.00%
 - Chopper Frequency
 - Detector Voltage (-0.100V to 0.100V)
- Set RFI Sensitivity
- Adjust internal barometer reading to match NIST-traceable barometer
- Adjust volume setting, if needed (1.50 L +/- 0.10 L)
- Print Technician Screen (attached)
- Print Calibration Factors (attached)
- Minimum Volume Test 1.5 L (1.3L to 1.7L)

As Left Conditions:

Barometric Pressure (in Hg): 29.83 Temperature (C): 21.4 Relative Humidity (%) 24

Linearity Test (attached)

0.020g/210L Mnf: <u>Airgas</u>	Lot: <u>AG217902</u>	Exp: <u>6/28/24</u>
0.040g/210L Mnf: <u>Airgas</u>	Lot: <u>AG217904</u>	Exp: <u>6/28/24</u>
0.080g/210L Mnf: <u>Airgas</u>	Lot: <u>AG310901</u>	Exp: <u>4/19/25</u>
0.100g/210L Mnf: <u>Airgas</u>	Lot: <u>AG211003</u>	Exp: <u>4/20/24</u>
0.150g/210L Mnf: <u>Airgas</u>	Lot: <u>AG134205</u>	Exp: <u>12/8/23</u>
0.300g/210L Mnf: <u>Airgas</u>	Lot: <u>AG217903</u>	Exp: <u>6/28/24</u>

Functional Tests (attached)

- Incomplete
 - Maximum Attempts Exceeded
 - Interference (Volatiles)
 - Invalid
 - Radio Frequency Interference (RFI)
- Acetone Mnf: Lifeloc Lot: 22HP4120 Level: 0.100 Exp: 8/24/24
Simulator S/N: DR4176 Verification Performed?
- Isopropanol Mnf: Lifeloc Lot: 22HP4120 Level: 0.100 Exp: 8/24/24
Simulator S/N: DR4177 Verification Performed?
- Neg. Control Mnf: Lifeloc Lot: 22HP4120 Level: 0.100 Exp: 8/24/24
Simulator S/N: DR6267 Verification Performed?

Calibrated External Thermometer:

Thermometer S/N: 210958384 Calibration Due Date: 11/9/23

- Verify stylus is calibrated
 - Reset Options
 - Clear Memory
 - Verification of Calibration (VOC)
 - Non-Drinking Subject Test
 - Upload
 - Remove Tank Prior to Shutdown
- Calibration End Date: 11/2/23

DIAGNOSTIC RECORD

Alaska Department of Public Safety
DATAMASTER dmt: 100673

Date: 10/31/2023
Time: 09:40:56

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
(44.0°C - 52.0°C)
Breath Tube = 46.5°C PASSED
(38.0°C - 50.0°C)

SETTINGS

Lamp Voltage = 1.70 V PASSED
(1.02 V - 2.38 V)
Cooler Voltage = 1.66 V PASSED
(1.00 V - 2.32 V)
Bias Voltage = 80 V PASSED
(48 V - 112 V)
Chopper Freq = 514 Hz PASSED
(475 Hz - 575 Hz)
Barometer = 29.9 in

PUMP INFO

Flow Rate = 4.443 L/M PASSED
(3.500 L/M - 6.500 L/M)

DETECTOR INFO

PUMP ON PASSED
(0.002 V <= 0.010 V)
PUMP OFF PASSED
(0.002 V <= 0.010 V)

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

Xq = 0.081 0.53% (0.00% - 4.00%) PASSED

Technician Screen

DATAMASTER dmt: 100673

Date: 10/31/2023
Time: 9:47 AM

Temperatures (°C)

Sample Cell: 48.9°C
Breath Tube: 45.8°C

Current Barometer 29.8 in
Volume (Ltr) 1.47

Settings

Lamp: 1.70 V
Bias: 80 V
Cooler: 1.66 V
Chopper: 516 Hz

Voltages (V)

Flow: 1.05 V
Detector: -0.075 V

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100673

Date: 07/18/2023
Time: 08:46:39

OPERATOR NAME:
TATIANA MALNOSKE
CERTIFICATE #: 10437

Ca = 0.1000
CAL = 0.989768 0.800 <= CAL < 1.200
b1 = 0.0011 0.0000 <= b1 < 0.0040
b2 = 0.0061 0.0010 <= b2 < 0.0100
b3 = 0.0010 0.0000 <= b3 < 0.0040
Xq = 0.0816 0.0500 <= Xq < 0.2000
a21 = 1.201042 1.050 <= a21 < 1.300
a31 = 0.440866 0.300 <= a31 < 0.800

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100673

Date: 10/31/2023
Time: 09:50:57

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX


OPERATOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436

DEPT/AGENCY:
MINIMUM VOLUME TEST
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.85: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK	PASSED	09:53
BLANK TEST	0.000	09:53
INTERNAL STANDARD	VERIFIED	09:53
EXTERNAL STANDARD	0.080	09:53
BLANK TEST	0.000	09:54
SUBJECT SAMPLE (Vol=1.50L)	0.000	09:55
BLANK TEST	0.000	09:56
EXTERNAL STANDARD	0.079	09:56
BLANK TEST	0.000	09:57

CVS



LINEARITY CHECK


 Alaska Department of Public Safety
 DATAMASTER dmt: 100673

Date: 10/31/2023
 Time: 12:36:13

SUPERVISOR NAME:
 CARLIE GLAISTER
 OPERATOR'S #: 10436

TARGET	AVERAGE	STD. DEV.
0.020	0.0200	0.0000
0.040	0.0398	0.0004
0.080	0.0794	0.0005
0.100	0.0998	0.0004
0.150	0.1496	0.0005
0.299	0.3006	0.0005

SLOPE = 1.006241
 R² = 0.999978



CKG

LINEARITY CHECK: Control Test 1

Alaska Department of Public Safety
 DATAMASTER dmt: 100673






Date: 10/31/2023
 Time: 12:36:15

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.020
 TARGET AT 29.87: 0.020
 LOT #: AG217902
 EXPIRATION: 06/28/2024
 TANK PRESSURE: 833 psi

BLANK TEST	0.000	12:37
INTERNAL STANDARD	VERIFIED	12:37
EXTERNAL STANDARD	0.020	12:37
BLANK TEST	0.000	12:38
EXTERNAL STANDARD	0.020	12:38
BLANK TEST	0.000	12:39
EXTERNAL STANDARD	0.020	12:40
BLANK TEST	0.000	12:40
EXTERNAL STANDARD	0.020	12:41
BLANK TEST	0.000	12:41
EXTERNAL STANDARD	0.020	12:42
BLANK TEST	0.000	12:43

Average = 0.0200
 Std Dev = 0.0000



LINEARITY CHECK: Control Test 2

Alaska Department of Public Safety
 DATAMASTER dmt: 100673






Date: 10/31/2023
 Time: 12:44:50

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.040
 TARGET AT 29.86: 0.040
 LOT #: AG217904
 EXPIRATION: 06/28/2024
 TANK PRESSURE: 967 psi

BLANK TEST	0.000	12:46
INTERNAL STANDARD	VERIFIED	12:46
EXTERNAL STANDARD	0.040	12:46
BLANK TEST	0.000	12:47
EXTERNAL STANDARD	0.040	12:47
BLANK TEST	0.000	12:48
EXTERNAL STANDARD	0.040	12:48
BLANK TEST	0.000	12:49
EXTERNAL STANDARD	0.040	12:49
BLANK TEST	0.000	12:50
EXTERNAL STANDARD	0.039	12:50
BLANK TEST	0.000	12:51

Average = 0.0398
 Std Dev = 0.0004



LINEARITY CHECK: Control Test 3

Alaska Department of Public Safety
 DATAMASTER dmt: 100673






Date: 10/31/2023
 Time: 12:53:44

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 29.85: 0.080
 LOT #: AG310901
 EXPIRATION: 04/19/2025
 TANK PRESSURE: 669 psi

BLANK TEST	0.000	12:55
INTERNAL STANDARD	VERIFIED	12:55
EXTERNAL STANDARD	0.079	12:55
BLANK TEST	0.000	12:56
EXTERNAL STANDARD	0.080	12:56
BLANK TEST	0.000	12:57
EXTERNAL STANDARD	0.079	12:57
BLANK TEST	0.000	12:58
EXTERNAL STANDARD	0.080	12:58
BLANK TEST	0.000	12:59
EXTERNAL STANDARD	0.079	12:59
BLANK TEST	0.000	13:00

Average = 0.0794
 Std Dev = 0.0005



LINEARITY CHECK: Control Test 4

Alaska Department of Public Safety
 DATAMASTER dmt: 100673






Date: 10/31/2023
 Time: 13:01:18

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.100
 TARGET AT 29.85: 0.100
 LOT #: AG211003
 EXPIRATION: 04/20/2024
 TANK PRESSURE: 543 psi

BLANK TEST	0.000	13:02
INTERNAL STANDARD	VERIFIED	13:02
EXTERNAL STANDARD	0.100	13:02
BLANK TEST	0.000	13:03
EXTERNAL STANDARD	0.100	13:04
BLANK TEST	0.000	13:04
EXTERNAL STANDARD	0.100	13:05
BLANK TEST	0.000	13:06
EXTERNAL STANDARD	0.100	13:06
BLANK TEST	0.000	13:07
EXTERNAL STANDARD	0.099	13:07
BLANK TEST	0.000	13:08

Average = 0.0998
 Std Dev = 0.0004



LINEARITY CHECK: Control Test 5

Alaska Department of Public Safety
 DATAMASTER dmt: 100673






Date: 10/31/2023
 Time: 13:12:06

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.150
 TARGET AT 29.85: 0.150
 LOT #: AG134205
 EXPIRATION: 12/08/2023
 TANK PRESSURE: 841 psi

BLANK TEST	0.000	13:13
INTERNAL STANDARD	VERIFIED	13:13
EXTERNAL STANDARD	0.149	13:13
BLANK TEST	0.000	13:14
EXTERNAL STANDARD	0.149	13:14
BLANK TEST	0.000	13:15
EXTERNAL STANDARD	0.150	13:16
BLANK TEST	0.000	13:16
EXTERNAL STANDARD	0.150	13:17
BLANK TEST	0.000	13:18
EXTERNAL STANDARD	0.150	13:18
BLANK TEST	0.000	13:19

Average = 0.1496
 Std Dev = 0.0005



LINEARITY CHECK: Control Test 6

Alaska Department of Public Safety
 DATAMASTER dmt: 100673






Date: 10/31/2023
 Time: 13:19:53


EXTERNAL STANDARD INFORMATION

NOMINAL: 0.300
 TARGET AT 29.84: 0.299
 LOT #: AG217903
 EXPIRATION: 06/28/2024
 TANK PRESSURE: 198 psi

BLANK TEST	0.000	13:21
INTERNAL STANDARD	VERIFIED	13:21
EXTERNAL STANDARD	0.300	13:21
BLANK TEST	0.000	13:22
EXTERNAL STANDARD	0.300	13:22
BLANK TEST	0.000	13:23
EXTERNAL STANDARD	0.301	13:24
BLANK TEST	0.000	13:24
EXTERNAL STANDARD	0.301	13:25
BLANK TEST	0.000	13:26
EXTERNAL STANDARD	0.301	13:26
BLANK TEST	0.000	13:27

Average = 0.3006
 Std Dev = 0.0005





SUBJECT TEST

 Alaska Department of Public Safety
 DATAMASTER dmt: 100673

 Date: 10/31/2023
 Time: 13:35:48

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
 CARLIE GLAISTER
 OPERATOR'S #: 10436

DEPT/AGENCY:
 INCOMPLETE
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 29.83: 0.080
 LOT #: AG310901
 EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK PASSED 13:38
 0.000 13:38
BLANK TEST
INTERNAL STANDARD VERIFIED 13:38
 0.080 13:38
EXTERNAL STANDARD
 0.000 13:39
BLANK TEST
SUBJECT SAMPLE INCOMPLETE 13:41

CUG

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100673

Date: 10/31/2023
Time: 13:43:21

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436


DEPT/AGENCY:
MAXIMUM ATTEMPTS EXCEEDED
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.83: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK PASSED 13:44
BLANK TEST 0.000 13:44
INTERNAL STANDARD VERIFIED 13:44
EXTERNAL STANDARD 0.080 13:44
BLANK TEST 0.000 13:45
SUBJECT SAMPLE XXX 13:46

Maximum Attempts Exceeded

GLG



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100673

Date: 10/31/2023
Time: 13:49:14

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436


DEPT/AGENCY:
INVALID
CASE/REPORT: XXXX
TEST TYPE: TEST


EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.84: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK PASSED 13:51
BLANK TEST 0.000 13:51
INTERNAL STANDARD VERIFIED 13:51
EXTERNAL STANDARD 0.079 13:51
BLANK TEST 0.000 13:52
SUBJECT SAMPLE INVALID 13:52

OK





SUBJECT TEST

 Alaska Department of Public Safety
 DATAMASTER dmt: 100673

 Date: 10/31/2023
 Time: 13:59:25

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
 CARLIE GLAISTER
 OPERATOR'S #: 10436

DEPT/AGENCY:
 RFI

CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 29.83: 0.080
 LOT #: AG310901
 EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK
 PASSED 14:00
 0.000 14:01
INTERNAL STANDARD
 VERIFIED 14:01
 0.079 14:01
EXTERNAL STANDARD
 0.000 14:02
INTERNAL STANDARD
 0.000 14:02

RFI DETECTED

OKG

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100673

Date: 10/31/2023
Time: 14:19:00

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436

DEPT/AGENCY:
INTERFERENCE - ACETONE
CASE/REPORT: XXXX
TEST TYPE: TEST


EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.83: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK PASSED 14:20
BLANK TEST 0.000 14:20
INTERNAL STANDARD VERIFIED 14:20
Xg = 0.082 (0.21%)
EXTERNAL STANDARD 0.079 14:20
X[1] = 0.0793 (-0.0008) (0.0032)
BLANK TEST 0.000 14:21
SUBJECT SAMPLE XXX 14:23
X[1] = 0.1114 (-0.0236) (0.0015)

Interference Detected

Filter 2 average = -0.0006
Filter 3 average = -0.0024
CUG 11/2/23

CUG



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100673

Date: 10/31/2023
Time: 14:05:26

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436

DEPT/AGENCY:
INTERFERENCE - IPA
CASE/REPORT: XXXX
TEST TYPE: TEST


EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.83: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK	PASSED	14:06
BLANK TEST	0.000	14:06
INTERNAL STANDARD	VERIFIED	14:06
Xq = 0.082 (0.00%)		
EXTERNAL STANDARD	0.079	14:06
X[1] = 0.0795 (-0.0011) (0.0024)		
BLANK TEST	0.000	14:07
SUBJECT SAMPLE	XXX	14:09
X[1] = 0.1189 (-0.0377) (0.0022)		

Interference Detected

CKU



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100673

Date: 10/31/2023
Time: 14:30:49

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX


OPERATOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436


DEPT/AGENCY:
INTERFERENCE - NEGATIVE
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.83: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK	PASSED	14:31
BLANK TEST	0.000	14:32
INTERNAL STANDARD	VERIFIED	14:32
$Xq = 0.082 (0.11\%)$	0.079	14:32
EXTERNAL STANDARD	$X[1] = 0.0031 (0.0031)$	14:33
BLANK TEST	0.000	14:33
SUBJECT SAMPLE (Vol=2.04L)	0.093	14:34
$X[1] = 0.0932 (-0.0033) (0.0007)$	0.000	14:35
BLANK TEST	0.079	14:35
EXTERNAL STANDARD	$X[1] = 0.0793 (-0.0006) (0.0008)$	14:36
BLANK TEST	0.000	14:36

OKG





SUBJECT TEST

 Alaska Department of Public Safety
 DATAMASTER dmt: 100673

 Date: 10/31/2023
 Time: 14:54:24

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
 CARLIE GLAISTER
OPERATOR'S #: 10436

DEPT/AGENCY:
 NON-DRINKING SUBJECT TEST
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.83: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK	PASSED	14:56
BLANK TEST	0.000	14:56
INTERNAL STANDARD	VERIFIED	14:56
EXTERNAL STANDARD	0.080	14:57
BLANK TEST	0.000	14:57
SUBJECT SAMPLE (Vol=1.68L)	0.000	14:58
BLANK TEST	0.000	14:59
EXTERNAL STANDARD	0.080	14:59
BLANK TEST	0.000	15:00

DKG-77D