

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form - DMCAF2011R1		
	Issuing Authority - Nita Bolz, Scientific Director		
Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active	

DataMaster DMT Calibration Form

Instrument # 100387Date: 4/12/11

## 1. Wet Bath Simulator Solution

- Level 0.100
- Manufacturer Guth
- Lot 10260
- Expiration Date 10/05/12

## 2. Wet Bath Simulator

- Serial Number DR-4161
- Date Calibrated 8/23/10

## 3. Attach Calibration Factors

Breath Alcohol Analyst C. Bryant

COB/2

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dat: 100387  
-----

Date: 04/12/2011  
Time: 14:12:44

SUPERVISOR NAME:  
COLLEEN S BRYANT  
CERTIFICATE #: 6237

LOT #: GUTH100  
BOTTLE #: 10260  
EXPIRATION: 10/25/2012

Ca	=	0.1000		
CAL	=	0.972396	0.800	<= CAL < 1.200
b1	=	0.0001	0.0000	<= b1 < 0.0040
b2	=	0.0051	0.0010	<= b2 < 0.0100
b3	=	0.0000	0.0000	<= b3 < 0.0040
Xq	=	0.0874	0.0500	<= Xq < 0.2000
a21	=	1.166124	1.050	<= a21 < 1.300
a31	=	0.471408	0.300	<= a31 < 0.800

*CSB*

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100387\B-100387\_ECF-20140

2/19/2014