

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 2/13/2025

Instrument #: 100366

Instrument Location: Kodiak Police Department, Jail, Datamaster Room

Supervisor Name and Agency: Phillip Christman, Kodiak Police Department

Describe the Status Change or Issue with the Instrument:

Tank pressure was low (above 100, less than 150). Tank was changed. VOC ran and did an additional test due to having some status messages which were possibly from the low pressure in the tank. One error was a filter wheel error. But after the tank change, I submitted a sample without issue.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Not taken out of service. Just changed tank.

Out of Service Date/Time: [Click or tap here to enter text.](#)

For Use by SCDL

Additional Notes

Verified 2 standard out of range, 1 filter wheel error, and 1 pump error status messages since 12/1/24. None of these status messages appear to be a trend. Change of the external standard tank would be recommended troubleshooting for standard out of range messages which was completed by the agency on 02/13/2025. Instrument will be monitored for filter wheel issues and standard out of range messages. DJW 02/14/2025

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor