

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 10/17/2022

Instrument #: 100422

Instrument Location: 8517 20th St Joint Base Elemendorf Richardson

Supervisor Name and Agency: Scott Gardner/6222/JBR4

Describe the Status Change or Issue with the Instrument:

Multiple status messages rcvd. Filters won't zero, ambient and blank errors. Failed diagnostic test.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 11 Oct 2022 @ 0948

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor