

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 10/15/2019 Instrument Serial Number: 100396

Instrument Location: Glacier Bay National Park

Supervisor Name and Agency: Colleen O'Bryant, SCDL

Describe the Status Change or Issue with the Instrument:

On 10/08/2019 Breath Test Supervisor, Wendy Bredow was describing a breath test she had performed on 05/26/2019 that included having the subject provide multiple breath samples. After researching the instrument it was discovered that the breath test options for 100396 were set to 3 tests rather than the standard 1 test. Only one test was performed on 100396 between it's in service date of 08/01/2018 and when the options were reset to 1 test (10/15/2019).

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Timeline of events -

04/10/2018 Instrument Certified

06/05/2018 Instrument used for training at Mat-Su College (set to 3 tests for training)

08/01/2018 Instrument In Service at GBNP

05/19/2019 Subject Test with 3 samples provided

10/09/2019 SCDL Notified about incorrect number of tests settings

10/15/2019 Wendy Bredow reset test settings to 1 test and performed non-drinking subject test (attached)

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

SUBJECT TEST

Alaska Department of Public Safety
 DATAMASTER dmt: 100396

Date: 10/15/2019
 Time: 08:23:05

SUBJECT NAME:
 TEST TEST
 LICENSE #: 123456

OPERATOR NAME:
 WENDY BREDOW
 OPERATOR'S #: 7797

DEPT/AGENCY:
 GBNP

CASE/REPORT: 00000
 TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 29.27: 0.078
 LOT #: AG735001
 EXPIRATION: 12/16/2019

DIAGNOSTIC CHECK	PASSED	08:24
BLANK TEST	0.000	08:24
INTERNAL STANDARD	VERIFIED	08:24
EXTERNAL STANDARD	0.075	08:24
BLANK TEST	0.000	08:25
SUBJECT SAMPLE (Vol=3.03L)	0.000	08:25
BLANK TEST	0.000	08:26
EXTERNAL STANDARD	0.076	08:26
BLANK TEST	0.000	08:27

WB

