

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100402

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/19/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.86: 0.080
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 1056 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 15:20 |
| INTERNAL STANDARD | VERIFIED | 15:20 |
| EXTERNAL STANDARD | 0.077 | 15:20 |
| BLANK TEST | 0.000 | 15:21 |
| EXTERNAL STANDARD | 0.077 | 15:21 |
| BLANK TEST | 0.000 | 15:22 |
| EXTERNAL STANDARD | 0.077 | 15:23 |
| BLANK TEST | 0.000 | 15:23 |
| EXTERNAL STANDARD | 0.077 | 15:24 |
| BLANK TEST | 0.000 | 15:25 |
| EXTERNAL STANDARD | 0.077 | 15:25 |
| BLANK TEST | 0.000 | 15:26 |

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 47.5°C PASSED

PUMP INFO

Flow Rate = 4.860 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only : Instrument being removed from
service.
JMB 7/11/17