

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100358

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 11/13/2025

## External Standard Test Values

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C

PASSED

Breath Tube = 48.1°C

PASSED

RFI DETECTED

File 001-  
CW 11/21/25