

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 11/20/17 Instrument Serial Number: 1006965 ^{CRS 11/20/17}

Instrument Location: Wasilla PD

Supervisor Name and Agency: Colleen O'Bryant

Describe the Status Change or Issue with the Instrument:

Instrument won't turn on. Power or screen
issue.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 11/20/17 10:59PM

For Use by SCDL

Additional Notes

Ben Dudley emailed saying the instrument would not
power up. I visited WPD on 11/20/17 and tested the
instrument with a new powercard + APC, unsuccessfully.
I removed the instrument from service + transported
it back to SCDL
OB 11/20/17

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

O'Bryant, Colleen S (DPS)

From: Barnett, Brandi M (DPS)
Sent: Tuesday, November 21, 2017 11:13 AM
To: bdudley@ci.wasilla.ak.us
Cc: O'Bryant, Colleen S (DPS)
Subject: FW: DataMaster Problem

Ben,

I just left you a message but have changed my mind a little bit. I am going to send a new instrument out with Colleen tonight and she will leave it at the Palmer Post. I want you to try the power cord that comes with the new instrument on your current instrument and see if that works. If it does then we will take the new one back. If not then you will already have the new instrument.

I just want to verify that you have tried bypassing the APC unit and plugging straight into the wall. It could be a problem with the APC too. I am also going to send you a new APC unit with your instrument in case that is the problem. Let us know how it goes.

Brandi

From: Foster, Charles R (DPS)
Sent: Tuesday, November 21, 2017 8:06 AM
To: Barnett, Brandi M (DPS)
Subject: FW: DataMaster Problem

Can you contact Wasilla PD to troubleshoot this today?

From: Benjamin Dudley [<mailto:bdudley@ci.wasilla.ak.us>]
Sent: Monday, November 20, 2017 10:59 PM
To: Foster, Charles R (DPS)
Subject: DataMaster Problem

I know this is not the proper channel to report this issue but I cannot find the general email address. I apologize in advance.

Our DataMaster at Wasilla PD does not seem to power up. I've plugged it into a couple different power outlets and nothing happens, just a blank screen and no noises.

Any suggestions?

Ben Dudley
Wasilla PD
907-352-5408