

**In Field Instrument Review – 04/11/2019 to 05/30/2019**

**Prepared by Colleen O’Bryant**

100352 (Anchorage Airport Police) – 1/14 Pump Error – No action taken

100355 (Seward PD) – 3/16 RFI, 1/16 Pump Error, 1/16 Internal Standard Error - Removed from service 06/13/2019

100357 (Dillingham PD) – 1/9 Internal Standard Error – Monitor instrument

100364 (Kodiak PD) – 2/12 RFI – No action taken

100365 (North Pole PD) – 2/9 Internal Standard Error – Removed from service 5/23/2019

100374 (Anchorage PD) – 1/130 Tank Interference - No action taken

100378 (AST/Fairbanks) – 1/43 Pump Error - No action taken

100379 (Wasilla PD) – 2/44 RFI - No action taken

100382 (UAF PD) – 1/9 Tank Interference - No action taken

100391 (AST/Soldotna) – 1/15 RFI - No action taken

100397 (NSB PD) – 1/7 Pump Error - No action taken

100402 (Seldovia PD) – 1/19 RFI - No action taken

100403 (Metlakatla PD) – 1/5 RFI - No action taken

100413 (Homer PD) – 1/13 RFI - No action taken

100416 (Fairbanks PD) – 1/35 RFI, 1/35 Tank Interference - No action taken

100418 (Kenai PD) – 1/22 Ambient Fail, 4/22 Diagnostic Fail –Breath tube temperature low –  
BTS reconnected breath hose – Monitor instrument.

100423 (Ft. WW) – 2/11 Pump Error - No action taken

100668 (Palmer PD) – 3/34 RFI - No action taken

100678 (AST/Mat-Su West) – 1/14 RFI - No action taken

100687 (AST/Soldotna) – 1/21 RFI, 1/21 Pump Error, 3/21 Tank Interference – Removed from service 5/15/2019

100690 (Unalaska DPS) – 1/7 Internal Standard Error – Monitor instrument

100691 (Valdez PD) – 1/2 Gas Flow Error - No action taken

100692 (AST/Girdwood) – 3/6 RFI - No action taken

100694 (Anchorage PD) – 1/95 Pump Error - No action taken

This review does not include data from the following instruments: 100362 (Fort Yukon PD), 100377 (AST/St. Mary's), 100424 (Eielson AFB), 100675 (AST/Angoon).



