

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 1/8/2025

Instrument #: 100415

Instrument Location: Anchorage Jail – South OUI processing room

Supervisor Name and Agency: Jason Cusack – Anchorage Police

Describe the Status Change or Issue with the Instrument:

On 1/3/2025 I was asked to check on instrument 100415 as the crime lab could not communicate with it. When I checked later that evening, I discovered it had not passed a VOC and when I attempted to run one it had several fatal errors. I blocked it off so no one would attempt to use it until I could remove it on 1/8/2025.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 1/8/2025 0740

For Use by SCDL

Additional Notes

Ticket db caused fatal errors. MNC 1/8/2025

It was noticed that an older version (4.0) of this form was originally submitted, all the information was transferred to the current version (6.0). See page 2 for the original. MNC 1/10/2025

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 1/8/2025

Instrument #: 100415

Instrument Location: Anchorage Jail – South OUI processing room

Supervisor Name and Agency: Jason Cusack – Anchorage Police

Describe the Status Change or Issue with the Instrument:

On 1/3/2025 I was asked to check on instrument 100415 as the crime lab could not communicate with it. When I checked later that evening, I discovered it had not passed a VOC and when I attempted to run one it had several fatal errors. I blocked it off so no one would attempt to use it until I could remove it on 1/8/2025.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 1/8/2025 0740

For Use by SCDL

Additional Notes

Ticket db caused fatal errors. MNC 1/8/2025

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor