

# Alaska Scientific Crime Detection Laboratory

## DataMaster Evaluation Form

Issued: 04/03/2018  
Effective: 04/04/2018

Version: DMEF2018 R0  
Status: Active

Instrument # 100673

Date: 09/09/20

Analyst: Derck Walton

1. Reason for instrument evaluation

Refurbished by factory for blank errors during attempted calibration

2. Visual Inspection

Clean

Replace missing screws

Cermetex Box is secure

Ensure modem cable present

Gas Valve upright or secured to wall

All tubing and connections are secure

3. O-ring inspection and ensure 2 O-rings in replacement bag

4. Troubleshooting/Repair

Controller board modification present. Sample chamber mirror blocks are not taped. Added o-ring.

5. VOC (attached)

6. Internal Standard value  $\leq 3.00\%$  Yes  No

7. Verify software version

Return to Factory for Repair

Ready for Calibration

Ready for Certification

Proceed to Close Out

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100673

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/09/2020

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.02: 0.080  
LOT #: AG826101  
EXPIRATION: 09/18/2020  
TANK PRESSURE: 356 psi

BLANK TEST	0.000	16:02
INTERNAL STANDARD	✓ VERIFIED	16:02
Xq = 0.080 (1.09%)		
EXTERNAL STANDARD	0.080	16:02
X[1] = 0.0801 (-0.0003) (0.0004)		
BLANK TEST	0.000	16:03
EXTERNAL STANDARD	0.079	16:03
X[1] = 0.0794 (-0.0010) (0.0003)		
BLANK TEST	0.000	16:04
EXTERNAL STANDARD	0.079	16:04
X[1] = 0.0795 (-0.0013) (0.0004)		
BLANK TEST	0.000	16:05
EXTERNAL STANDARD	0.080	16:05
X[1] = 0.0798 (-0.0007) (0.0004)		
BLANK TEST	0.000	16:06
EXTERNAL STANDARD	0.080	16:07
X[1] = 0.0796 (-0.0012) (0.0007)		
BLANK TEST	0.000	16:08

Average = 0.0796  
Std Dev = 0.0005

## Diagnostic Check

VERSIONS ✓  
DMT: 3.02 ✓  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 46.9°C	PASSED
(38.0°C - 50.0°C)	

### SETTINGS

Lamp Voltage = 1.70 V	PASSED
(1.02 V - 2.38 V)	
Cooler Voltage = 1.71 V	PASSED
(1.03 V - 2.39 V)	
Bias Voltage = 80 V	PASSED
(48 V - 112 V)	
Chopper Freq = 510 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 30.0 in	

### PUMP INFO

Flow Rate = 4.179 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

### DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD ✓	PASSED
Xq = 0.080 0.33% (0.00% - 4.00%)	