

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.2

Effective: 1/17/2024

Date: 1/6/2025

Instrument #: 100403

Instrument Location: Wasilla PD

Supervisor Name and Agency: Derek Walton – SCDL working with Amanda Graham – Wasilla PD

Describe the Status Change or Issue with the Instrument:

I spoke to Amanda on the phone regarding the Wasilla PD instrument not polling. Amanda is an IT professional with Wasilla PD. She verified that the modem line was connected to the DataMaster and that she believed it to be functioning correctly. I requested that she turn off the instrument and then turn in back on after 5 seconds which she did. She notified me that the instrument rebooted normally, was able to clear the driver message, and print a copy of the most recent record with no issues. I told Amanda that I would call the instrument from the lab to see if the restart fixed the polling. Carlie Glaister was then able to retrieve files via ~~normal~~ polling on 01/06/205.
normal DJW 01/31/2025

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor