

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 12/18/19 Instrument Serial Number: 100361
Instrument Location: Patrol Room B.P.D.
Supervisor Name and Agency: Justin Ulak Bethel PD

Describe the Status Change or Issue with the Instrument:

Removing from service. 'Detector overflow' error message.
Being replaced.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 12/18/19, 1900 hours

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740