

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100350

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/01/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.59: 0.079
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 326 psi

BLANK TEST	0.000	09:52
INTERNAL STANDARD	VERIFIED	09:52
EXTERNAL STANDARD	0.077	09:52
BLANK TEST	0.000	09:53
EXTERNAL STANDARD	0.078	09:53
BLANK TEST	0.000	09:54
EXTERNAL STANDARD	0.079	09:55
BLANK TEST	0.000	09:55
EXTERNAL STANDARD	0.079	09:56
BLANK TEST	0.000	09:57
EXTERNAL STANDARD	0.078	09:57
BLANK TEST	0.000	09:58

Average = 0.0782
Std Dev = 0.0008

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 40.7°C PASSED

PUMP INFO

Flow Rate = 4.353 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only: Instrument being removed from
service. BMB 9/13/17