

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100684

Received From Delta AST

Date 04/03/18

Analyst Derek Walton / Brandi Barnett

1. Reason for instrument return:

Standard out of range, Internal Standard error, Gas cylinder leaking during tank change

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload Does not turn on. Cannot upload.

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs Does not turn on. Cannot upload.

From Intoximeters

7. Intoximeters Repair Paperwork (attached)