

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Effective: 11/10/2021

Version: 2.0

Date: 1-7-2022 Instrument #: 100366

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Nicholas Bone APD

Describe the Status Change or Issue with the Instrument: Failed VOC

internal standard error + pump error

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 4/6/2022 2100

For Use by SCDL	Additional Notes

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740