

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100359

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:08/03/2020

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 44.6°C PASSED

PUMP INFO

Flow Rate = 4.500 L/M PASSED

DETECTOR INFO

PUMP ON FAILED
PUMP OFF FAILED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

Diagnostic Check Failed

File only DM 09/15/20