

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 12/18/2019 Instrument Serial Number: 100347  
Instrument Location: Patrol Room, BPD  
Supervisor Name and Agency: Justin Ulak, Bethel PD

Describe the Status Change or Issue with the Instrument:

Instrument sent to replace another that was taken out of service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 12/18/19, 1905 hours

Out of Service Date/Time: \_\_\_\_\_

For Use by SCDL

Additional Notes

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740