

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100424

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/20/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.40: 0.079  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 755 psi

BLANK TEST	0.000	00:59
INTERNAL STANDARD	VERIFIED	00:59
EXTERNAL STANDARD	0.077	01:00
BLANK TEST	0.000	01:01
EXTERNAL STANDARD	0.077	01:01
BLANK TEST	0.000	01:02
EXTERNAL STANDARD	0.077	01:02
BLANK TEST	0.000	01:03
EXTERNAL STANDARD	0.077	01:03
BLANK TEST	0.000	01:04
EXTERNAL STANDARD	0.077	01:04
BLANK TEST	0.000	01:05

Average = 0.0770  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 47.6°C PASSED

### PUMP INFO

Flow Rate = 3.915 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD


PASSED

I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Charles R. Foster  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 7 day of Jun, 20 19

  
Notary Public  
My Commission Expires With Office



JW 05/29/19