	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form – DMCAF2011R1		
	Issuing Authority – Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100421

Date: 7/28/11


1. Wet Bath Simulator Solution

- a. Level 0.10
- b. Manufacturer Guth
- c. Lot 11050
- d. Expiration Date 4/6/13

2. Wet Bath Simulator

- a. Serial Number D124177
- b. Date Calibrated 12/16/10

3. Attach Calibration Factors

Breath Alcohol Analyst 
 Charles Foster

1/2 CR1

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100421

Date: 07/28/2011
Time: 14:46:11

SUPERVISOR NAME:
CHARLES FOSTER
CERTIFICATE #: 7734

LOT #: 11050
BOTTLE #: 705
EXPIRATION: 04/06/2013

Ca = 0.1000
CAL = 1.000419 0.800 <= CAL < 1.200
b1 = 0.0004 0.0000 <= b1 < 0.0040
b2 = 0.0027 0.0010 <= b2 < 0.0100
b3 = 0.0007 0.0000 <= b3 < 0.0040
Xq = 0.1013 0.0500 <= Xq < 0.2000
a21 = 1.117763 1.050 <= a21 < 1.300
a31 = 0.453616 0.300 <= a31 < 0.800

2/2 CTF

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100421\B-100421 ECF-20140

2/18/2014