

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100693

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/15/2013

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.35: 0.081
LOT #: 02012080A1
EXPIRATION: 03/01/2014

BLANK TEST	0.000	01:29
INTERNAL STANDARD	VERIFIED	01:30
EXTERNAL STANDARD	0.079	01:30
BLANK TEST	0.000	01:31
EXTERNAL STANDARD	0.079	01:31
BLANK TEST	0.000	01:32
EXTERNAL STANDARD	0.079	01:32
BLANK TEST	0.000	01:33
EXTERNAL STANDARD	0.080	01:33
BLANK TEST	0.000	01:34
EXTERNAL STANDARD	0.079	01:34
BLANK TEST	0.000	01:35

Average = 0.0792
Std Dev = 0.0004

Diagnostic Check

VERSIONS
DMT: 2.00
PIC: 3.03
Modem: 2.3
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
Breath Tube = 45.0°C PASSED

PUMP INFO

Flow Rate = 4.560 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO


Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

10/22/2013
10/22/13 NJB
2013

Subscribed and sworn before me this 22 day of Oct, 20 13


Nikki Roth, Notary Public
My Commission Expires With Office



NJB
10/17/13