

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 08/15/18 Instrument Serial Number: 100405

Instrument Location: Dillingham PD

Supervisor Name and Agency: Dan Pasquariello - Dillingham PD

Describe the Status Change or Issue with the Instrument:

Multiple interference detected status messages during BTD certification class.  
Successful VAC and non-drinking subject test next day (06/14/2018).

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: NA ~~\_\_\_\_\_~~ JMV  
08/15/18

For Use by SCDL

Additional Notes

No status messages since. No further action taken. JMV 08/15/2018

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740