

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 9/30/2024

Instrument #: B100346

Instrument Location: WildWood Pretrial

Supervisor Name and Agency: Chad Larsen, Kenai Police Department

Describe the Status Change or Issue with the Instrument:

The above listed Datamaster is now in good reliable working condition after removing the cold air that was blowing on the instrument as instructed by Megan Carlock, DPS.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 09/23/2024 / 1334 hours

Out of Service Date/Time: 09/23/2024 / 0937 hours

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor