

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 11/10/2021

Version: 2.0

Instrument #: 100345

Received From: Valdez PD

Date: August 15, 2022

Analyst: Tatiana Malnorko

Supervised by Carlie Glaister

Reason for instrument return:

High Prevalence of standard out of range
status messages.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)