

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100359

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:08/14/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.21: 0.081
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 1018 psi

BLANK TEST	0.000	14:43
INTERNAL STANDARD	VERIFIED	14:44
EXTERNAL STANDARD	0.079	14:44
BLANK TEST	0.000	14:45
EXTERNAL STANDARD	0.078	14:45
BLANK TEST	0.000	14:46
EXTERNAL STANDARD	0.079	14:46
BLANK TEST	0.000	14:47
EXTERNAL STANDARD	0.078	14:47
BLANK TEST	0.000	14:48

Detector Overflow

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.6°C	PASSED
Breath Tube = 47.0°C	PASSED

PUMP INFO

Flow Rate = 4.521 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------