

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Effective: 11/10/2021

Version: 2.0

Date: 5/6/2022 \_\_\_\_\_ Instrument #: 100669 \_\_\_\_\_

Instrument Location: AST/Soldotna \_\_\_\_\_

Supervisor Name and Agency: Carlie Glaister - SCDL \_\_\_\_\_

Describe the Status Change or Issue with the Instrument:

When reviewing status messages during IFIR it was noticed that AST Soldotna has a higher percentage of RFI status messages than average. An email was sent to the Breath Test Supervisor, Joseph Miller, to remind Breath Test Operators and other officers not to key radios within the vicinity of the DataMaster when it is in use.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA \_\_\_\_\_

Out of Service Date/Time: NA \_\_\_\_\_

**For Use by SCDL**

Additional Notes

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**

For questions contact the Breath Alcohol Program at 907-269-5740