

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 6/21/2024

Instrument #: 100361

Instrument Location: North Processing Room Anchorage Jail

Supervisor Name and Agency: Jason Cusack Anchorage Police

Describe the Status Change or Issue with the Instrument:

The plastic piece that holds the mouthpiece onto the breath hose separated and was not realized until it was somewhere in the bottom of a large trash can. I conducted a VOC before removing the instrument from service because it will not hold a mouth piece onto the breath hose.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 6/21/2024 0100

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor