

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 7/7/17 Instrument Serial Number: 100382

Instrument Location: UAF PD

Supervisor Name and Agency: Brandi Barnett SCDL

Describe the Status Change or Issue with the Instrument:

See attached documentation email.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

Barnett, Brandi M (DPS)

From: Barnett, Brandi M (DPS)
Sent: Friday, July 07, 2017 7:53 AM
To: sabeaudreault@alaska.edu
Cc: Foster, Charles R (DPS); O'Bryant, Colleen S (DPS)
Subject: FW: Datamaster

Sam,

I have looked at your tests from 7/1/2017 and the graph shows the alcohol line with a negative slope in both cases. Due to this the instrument should have triggered an invalid. With only tests from the one subject it is not possible for me to tell if the issue was with the subject or the instrument. Is the officer certain that there was no possibility for mouth alcohol? Fatal errors happen occasionally and a reboot usually fixes those. I would like to just keep an eye on the instrument and see if there are any other issues. Please let us know as soon as possible if you continue to get unusual status messages or more invalid status messages. We will keep an eye on it as well. Also, I will document this in the instrument file.

Brandi

From: DPS SCDL Toxicology (DPS sponsored)
Sent: Thursday, July 06, 2017 9:51 AM
To: O'Bryant, Colleen S (DPS); Barnett, Brandi M (DPS); Foster, Charles R (DPS)
Subject: FW: Datamaster

From: Samuel Beaudreault [<mailto:sabeaudreault@alaska.edu>]
Sent: Wednesday, July 05, 2017 11:00 PM
To: DPS SCDL Toxicology (DPS sponsored)
Subject: Re: Datamaster

Did another test today and it worked fine.

On Sat, Jul 1, 2017 at 1:36 AM, Samuel Beaudreault <sabeaudreault@alaska.edu> wrote:
Our Datamaster is still acting up.

One of our officers had a DUI tonight and reported it had status messages of Fatal Error, then two Invalid Samples after that, while testing the same person.

--
Officer Sam Beaudreault
UAF Police Department
ph. [\(907\)474-7721](tel:9074747721)
fax [\(907\)474-5555](tel:9074745555)

--
Officer Sam Beaudreault
UAF Police Department
ph. (907)474-7721
fax (907)474-5555