

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 1/13/2017  
Effective: 1/17/2017

Version: DMCIF 2017 R0  
Status: Active

Instrument S/N 100678

Received From AST/Anchor Point

Date 07/30/2018

Analyst Derek Walter

1. Reason for instrument return:

Instrument freezing in the field

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload  Won't upload via modem line, uploaded via USB

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

### From Intoximeters

7. Intoximeters Repair Paperwork (attached)