



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 1234
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 02511080A3
Expiration: 2/1/2013

0.080 BAC (For use with breath alcohol testing instruments)

Contents: 105 Liters @ 1025 psig 70°F (21°C)


Component:	Concentration:	Accuracy:	Method:
Ethanol	208.4 ppm	+/- 0.002 or 2% BAC whichever is greater	NDIR
Nitrogen	Balance		

Specialty Gas Lab Manager

1-28-2011

Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com



SUPERVISOR TEST

Alaska Department of Public Safety
 DATAMASTER dmt: 100359

Date: 03/15/2011
 Time: 09:35:23

SUPERVISOR NAME:
 COLLEEN S OBRYANT
 OPERATOR'S #: 6237

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 29.20: 0.078

BLANK TEST	0.000	09:36
INTERNAL STANDARD	VERIFIED	09:36
EXTERNAL STANDARD	0.080	09:36
BLANK TEST	0.000	09:37
EXTERNAL STANDARD	0.079	09:38
BLANK TEST	0.000	09:38
EXTERNAL STANDARD	0.078	09:39
BLANK TEST	0.000	09:40
EXTERNAL STANDARD	0.079	09:40
BLANK TEST	0.000	09:41
EXTERNAL STANDARD	0.078	09:41
BLANK TEST	0.000	09:42

Average = 0.0788
 Std Dev = 0.0008

*Lot 02511080A3
 Exp 3/1/13*

