

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100685

Received From Intoximeters

Date 02/03/21

Analyst Dereu Walton

1. Reason for instrument return:

Refurbished by the manufacturer

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

*Instrument was not in the field
JMW 02/03/21*

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)

SRO Number: SRO-043152

Received Date: 11/25/2020

Completion Date: 01/06/2021

SRO Type: REPAIR

SRO Description: DMT Preventative Maintenance

Customer ID: C000003651

Customer PO: 112520-Barnett

Contact: Derek Walton

Email: derek.walton@alaska.gov

Ship To:

Alaska Department of Public Safety
Procurement & Supply Section
4805 Dr. MLK Jr. Avenue
Anchorage AK 99507

Units on SRO

1 LEGDMT LEGACY DMT

Service Line: 1	**All Instruments are Calibrated to Factory Specifications**
Unit Type:	(LEGDMT) LEGACY DMT
Serial Number:	50100685
Reason:	Preventive Maintenance
	planned
Reason Notes:	For Preventive Maintenance.
Resolution:	Update
	DMTVR3
Resolution Notes:	Installed TB-30 DMT Digital Pots Mod on Control Board. Cleaned & adjusted filter wheel solenoid. Updated filter wheel solenoid mounting screws. Updated stepper motor adjustment lock down screw. Replaced filter wheel motor pinion. Cleaned, lubricated, & adjusted filter wheel moving parts. Applied epoxy to stepper motor wires. Removed, inspected, and applied metallic seal tape to mirror blocks of Sample Chamber. Replaced discolored tubing and inspected 5-Way Valve. Optimized detector position. Adjusted Tech Screen settings. Performed Calibration/Adjustment, Wet and Dry Gas Accuracy Checks, Breath Tests, and outgoing Diagnostics Test.

Parts:	Qty	Part Number	Part Description
	1	90-8000-00	CONTRACT PREV MAINT DMT
	1	001199-00	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN

Order Packing Slip Report

Intoximeters
Experience • Service • Integrity

1/14/2021 2:06:35 PM

Packing Slip: 219057

Pack Date: 1/14/2021

SRO: SRO-043152



From: Intoximeters, Inc
2081 Craig Road
St Louis MO 63146
USA

Customer: C000003651

Ship To: 2

Alaska Department of Public Safety
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Alaska Department of Public Safety
Procurement & Supply Section
4805 Dr. MLK Jr. Avenue
Anchorage AK 99507
USA

Customer PO: 112520-Barnett

Ship Via: FDX

FED EX

Weight: 37.00

Delivery Terms : 2D

FOB : ORIGIN

Packages: 1

Customer Ship Account :

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity	
1	LEGDMT -	LEGACY DMT			50100685	
1	90-8000-00	EA		1.00	0.00	
	CONTRACT PREV MAINT DMT					
2	001199-00	EA		1.00	0.00	
	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN					

Rec'd 02/03/21
Jm