

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100667

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 10/24/2021

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 27.71: 0.074  
LOT #: AG113402  
EXPIRATION: 05/14/2023  
TANK PRESSURE: 1230 psi

BLANK TEST                      0.000    12:02  
INTERNAL STANDARD            ERROR    12:02

Internal Standard Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C            PASSED  
Breath Tube       = 45.4°C           PASSED

### PUMP INFO

Flow Rate = 4.244 L/M               PASSED

### DETECTOR INFO

PUMP ON                               PASSED  
PUMP OFF                               PASSED

### FILTER INFO

Filter 1                                PASSED  
Filter 2                                PASSED  
Filter 3                                PASSED

INTERNAL STANDARD                   PASSED