

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Acceptance Form - DMAF2011R0		
	Issuing Authority - Nita Bolz, Scientific Director		
Approval date: 03/01/11	Effective date: 03/01/11	Archive date: Active	

DataMaster DMT Acceptance Form

Instrument # 100422

Date: 8/19/11

Analyst: Colleen O'Bryant

1. Check in instrument in to JusticeTrax Breath Alcohol Database
2. Visual Inspection
3. Update software
4. Print calibration factors (attached)
5. NPAS calibration paperwork (attached)
6. Barcode Instrument
7. State of Alaska Property ID# _____

*Check in not completed. Software not updated.
NPAS paperwork did not come with instrument
when delivered from JBEK.*

*Returned to NPAS - interference with external
standard.*

*Replaced w/ new sticker + updated instrument
serial number
08/19/11*



old sn

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100422

Date: 08/15/2011
Time: 18:31:02

SUPERVISOR NAME:
COLLEEN S OBRYANT
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 11050
EXPIRATION: 04/06/2013

Ca = 0.1000
CAL = 0.991964 0.800 <= CAL < 1.200
b1 = 0.0001 0.0000 <= b1 < 0.0040
b2 = 0.0056 0.0010 <= b2 < 0.0100
b3 = 0.0002 0.0000 <= b3 < 0.0040
Xq = 0.0960 0.0500 <= Xq < 0.2000
a21 = 1.229637 1.050 <= a21 < 1.300
a31 = 0.414175 0.300 <= a31 < 0.800

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100422\B-100422 ECF-20140

7/9/2014