

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 11/26/2022

Instrument #: 100385

Instrument Location: 1979 Peger Road, Fairbanks, AK 99701

Supervisor Name and Agency: Trooper James W. Thomas III; Alaska State Troopers

Describe the Status Change or Issue with the Instrument:

Unexpected error in software. System.NullReferenceException: NulReferenceException at NPAS.  
DMT will not run nor can the setup menu be accessed. This error does not appear in the manual.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 11/26/2022 0045 hours

-----  
**For Use by SCDL**

Additional Notes

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor