

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100384

Received From Intoximeters

Date 02/03/21

Analyst Derek Walton

1. Reason for instrument return:

Refurbished by manufacturer for sample
chamber temp = 0

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

Instrument was not
in the field
DW 02/03/21

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)

SRO Number: SRO-043154

Received Date: 11/25/2020

Completion Date: 01/14/2021

SRO Type: REPAIR

SRO Description: DMT Preventative Maintenance

Customer ID: C000003651

Customer PO: 112520-Barnett

Contact: Derek Walton

Email: derek.walton@alaska.gov

Ship To:

Alaska Department of Public Safety
Procurement & Supply Section
4805 Dr. MLK Jr. Avenue
Anchorage AK 99507

Units on SRO

1 LEGDMT LEGACY DMT

Service Line: 1	**All Instruments are Calibrated to Factory Specifications**
Unit Type:	(LEGDMT) LEGACY DMT
Serial Number:	50100384
Reason:	Preventive Maintenance
	planned
Reason Notes:	For Preventative Maintenance
Resolution:	Prev M
	planned
Resolution Notes:	<p>90-8000-00 Prev Maint Contract p/n Replaced U28 Dc DC converter Installed digital pot mod on the control board Adjusted tech screen settings and RFI sensitivity Added seal tape and new torx screws to sample chamber. Mounted dual check valve to enclosure wall Replaced breath tube o ring Adjusted Time and Date Adjusted and accuracy checked</p> <p>Parts 001199-00 3.3volt regulator 59-1030-00 DC DC converter 59-0131-00 2pos. Terminal connector</p> <p>Calibration solution (0.200) Lot#20160 Bottle#1091 Dry gas: AG021605, Tank 011, Exp. 03AUG2022</p>

Parts:	<u>Qty</u>	<u>Part Number</u>	<u>Part Description</u>
	1	59-1030-00	DMT ELEC CONVERTER DC-DC 01719
	1	59-0131-00	DMT ELEC SCREW TERMINAL 2 POS 36596
	1	001199-00	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN
	1	90-8000-00	CONTRACT PREV MAINT DMT

Order Packing Slip Report

Intoximeters
Experience · Service · Integrity

1/25/2021 9:48:55 AM

Packing Slip: 219756

Pack Date: 1/25/2021

SRO: SRO-043154



From: Intoximeters, Inc
2081 Craig Road
St Louis MO 63146
USA

Customer: C000003651

Ship To: 2

Alaska Department of Public Safety
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Alaska Department of Public Safety
Procurement & Supply Section
4805 Dr. MLK Jr. Avenue
Anchorage AK 99507
USA

Customer PO: 112520-Barnett

Ship Via: FDX

FED EX

Weight: 38.00

Delivery Terms : 2D

FOB : ORIGIN

Packages: 1

Customer Ship Account :

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100384
1	59-1030-00	EA	1.00	1.00	0.00
	DMT ELEC CONVERTER DC-DC 01719				
2	59-0131-00	EA	1.00	1.00	0.00
	DMT ELEC SCREW TERMINAL 2 POS 36596				
3	001199-00	EA	1.00	1.00	0.00
	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN				
4	90-8000-00	EA	1.00	1.00	0.00
	CONTRACT PREV MAINT DMT				

Rec'd 02/03/21
DMV