

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 6/16/21 Instrument Serial Number: 100397
Instrument Location: 1068 KIOGAK ST, UTQIAGVIK, AK, 99723 / PD SQUADROOM
Supervisor Name and Agency: KEVIN NELSON / NSBPD

Describe the Status Change or Issue with the Instrument:

POSSIBLE MODEM ISSUE. PHONE LINE TESTED FOR DMT. TEST WAS GOOD
SENDING BACK TO CRIME LAB.

REPLACEMENT DMT RECEIVED TODAY

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 6/16/21 12:07 PM

For Use by SCDL

Additional Notes

Removed from service because preventive maintenance was due. DJW 06/16/2021

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740