

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 1/13/2017  
Effective: 1/17/2017

Version: DMCIF 2017 R0  
Status: Active

Instrument S/N 100417

Received From Intoximeters

Date 04/10/20

Analyst Derech walton

1. Reason for instrument return:

Instrument refurbished by factory, such back error  
during laboratory testing

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

Instrument was not in the field DW 04/10/20

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

### From Intoximeters

7. Intoximeters Repair Paperwork (attached)

**SRO Number:** SRO-039492

**Received Date:** 03/11/2020

**Completion Date:** 03/25/2020

**SRO Type:** REPAIR

**SRO Description:** DMT Repair/Prev Maint

**Customer ID:** C000003651

**Customer PO:** 20200108DJW

**Contact:** Derek Walton

**Email:** derek.walton@alaska.gov

Ship To:

Alaska Scientific Crime Detection Laboratory  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

Units on SRO

1	LEGDMT	LEGACY DMT
---	--------	------------

<b>Service Line:</b> 1	<b>**All Instruments are Calibrated to Factory Specifications**</b>
<b>Unit Type:</b>	(LEGDMT) LEGACY DMT
<b>Serial Number:</b>	50100417
<b>Reason:</b>	Preventive Maintenance planned
<b>Reason Notes:</b>	Preventive Maintenance
<b>Resolution:</b>	Prev M planned
<b>Resolution Notes:</b>	Installed digital pot Mod to Control Board Rebuilt filter assembly and added epoxy to motor wires Optimized detector Applied metal seal tape to the sample chamber mirror blocks. Adjust tech screen settings and RFI sensitivity Added o ring to end of breath tube Calibrated Ran wet and dry accuracy checks  Parts 001199-00 3.3volt regulator  Calibration solution (0.200) Lot# 19200 Bottle# 388 Dry gas: AG909102, Tank 044, Exp. 04/01/2021

**Parts:** Qty      Part Number      Part Description

Qty	Part Number	Part Description
1	001199-00	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN
<b>Reason:</b>	DMT Status Messages Suckbk	
<b>Reason Notes:</b>	Customer Reported Issue: persistent suck back error	
<b>Resolution:</b>	Did not Did not	
<b>Resolution Notes:</b>	Unable to verify complaint.	
1	90-8000-00	CONTRACT PREV MAINT DMT